

KI0443 HPLC Analysis Report

Sample: Axitinib

Sample ID: KI0443

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 35-55% Buffer B in 20min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT(min)	Area%	Area(mAU*s)	Height(mAU)
	1	8.797	0.7766	135992	16538
	2	9.518	98.65	17276200	1541744
	3	13.778	0.5716	100092	10418
Total			100	17512284	1568700