

KI0192 HPLC Analysis Report

Sample: CAL-101

Sample ID: KI0192

Buffer A: 0.1% TFA in 100% water (v/v)


Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)

Gradient: 39-59% Buffer B in 20min

Flow: 1ml/min Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
	1	8.763	0.03026	10160	16729
	2	8.882	99.89	33536592	4705195
	3	9.230	0.07806	26205	11618
Total			100	33572957	4733542