

HPLC Analysis Report

Sample: **Exemestane**

CAS No.: 107868-30-4

Synonym: 6-Methylideneandrosta-1,4-diene-3,17-dione

Method Info: 10%-90% Buffer B 0-5 MIN

Buffer A: 100% water


Buffer B: 100% acetonitrile

Column: XBridge RP18, 3*100mm, 3.5um

Flow Rate: 1mL/min

Wavelength: 254 nm

Date: May. 29, 2014



Signal Peak#	RT(min)	Width(min)	Area(mAU*s)	Height(mAU)	Area%
1	3.785	0.1034	5.6985	0.74226	0.2004
2	3.983	0.0722	2834.04443	609.85645	99.7996
Total			2839.73528	610.59871	100.0000

Analysis Result:

Purity: 99.8%