

KI0301 HPLC Analysis Report

Sample: Imatinib

Sample ID: KI0301

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 10-70% Buffer B in 20min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
	1	11.868	0.3982	120614	19949
	2	13.028	99.6	30172765	1723134
Total			100	30293379	1743083