

**Ostarine, MK-2866 HPLC Analysis Report**

Sample: **Ostarine, MK-2866**

Sample ID: **KI0132**

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 69-89% Buffer B in 20min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6\*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
	1	10.047	98.09	15329781	1289253
	2	10.773	0.9004	140711	4555
	3	11.223	0.8556	133716	9474
	4	11.658	0.1537	24028	1965
<b>Total</b>			<b>100</b>	<b>15628236</b>	<b>1305247</b>