

KI0055 HPLC Analysis Report

Sample: SNS-032

Sample ID: **KI0055**

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 10-100% Buffer B in 30min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
	1	17.037	0.3947	158868	23408
	2	17.865	99.24	39942003	4411285
	3	19.367	0.3665	147497	23615
Total			100	40248368	4458308