

KI0608 HPLC Analysis Report

Sample:TAK-733

Sample ID: **KI0608**

Buffer A:0.1% TFA in 100% water(v/v)


Buffer B:0.1% TFA in 80% acetonitrile + 20% water(v/v)

Gradient:10-100% Buffer B in 30min and 100% Buffer B in 10min

Flow:1ml/min Wavelength:220nm

Column:Venusil C18 5um4.6*250mm

Chromatogram



Signal	Peak#	RT(min)	Area%	Area(mAU*s)	Height (mAU)
	1	21.453	99.58	11217643	1829873
	2	22.250	0.4174	47013	7442
Total			100	11264656	1837315