

KI0617 HPLC Analysis Report

Sample: Torin2

Sample ID: KI0617

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 10-100% Buffer B in 30min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
	1	15.007	0.3212	18315	1642
	2	15.356	99.25	5660136	889732
	3	16.643	0.4289	24461	4107
Total			100	5702912	895481