

Vorinostat HPLC Analysis Report

Sample: Vorinostat

Sample ID: **KI0176**

Buffer A: 0.1% TFA in 100% water (v/v)


Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)

Gradient: 10-100% Buffer B in 30min

Flow: 1ml/min Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



| Signal | Peak# | RT (min) | Area% | Area (mAU*s) | Height (mAU) |
|--------|-------|----------|--------|--------------|--------------|
| | 1 | 15.423 | 99.85 | 30565466 | 2767108 |
| | 2 | 15.788 | 0.1518 | 46466 | 10120 |
| Total | | | 100 | 30611932 | 2777228 |