

KI0043 HPLC Analysis Report

Sample: YZ7632

Sample ID: KI0043

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 10-70% Buffer B in 20min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
	1	8.260	0.6232	4826	3830
	2	8.395	99.24	768582	96377
	3	8.557	0.1345	1042	695
Total			100	774450	100902